

dead time correction (in radioanalytical chemistry)

Correction to be applied to the observed number of pulses in order to take into account the number of pulses lost during the resolving or *dead time*.

1994, 66, 2518