

efficiency of atomization (in analytical flame spectroscopy)

The ratio of the amount of analyte that passes through the flame cross-section at the observation height, as free neutral (or ionized) atoms, to the amount of analyte aspirated.

O.B. 124