

geometry factor (in radioanalytical chemistry)

The average solid angle in steradians at source subtended by the aperture or sensitive volume of the detector, divided by 4π .

1994, 66, 2520