

shut-down state (in analysis)

The condition of an instrument when it is switched off to conserve energy or reagents or to protect working parts. This term is of particular importance in clinical chemistry.

1989, *61*, 1664